

ARTICLE

# Real-Time Disease Detection: Building IoT Remote Monitoring Systems with Neural Networks

Mohammed Al-Qahtani, Ahmed Al-Harbi, Fatimah Al-Otaibi, and Aisha Al-Shehri\*

Department of Computer Science, College of Computer and Information Sciences, King Saud University, P.O. Box 51178, Riyadh 11543, Saudi Arabia

\*Corresponding author: [Aishaal.she86ri@ksu.edu.sa](mailto:Aishaal.she86ri@ksu.edu.sa)

(Received: 14 October 2024; Revised: 04 February 2025; Accepted: 05 March 2025; Published: 23 March 2025)

## Abstract

The integration of Internet of Things (IoT) technologies with artificial intelligence is revolutionizing healthcare by enabling real-time disease detection and patient monitoring. This study presents the design and implementation of an IoT-based remote health monitoring system that utilizes neural networks for early disease diagnosis. Biomedical sensors continuously capture vital parameters such as heart rate, body temperature, oxygen saturation, and electrocardiogram signals, which are transmitted via IoT gateways to cloud servers for intelligent analysis. A deep neural network (DNN) model processes the sensor data to identify abnormal patterns associated with potential diseases, achieving high accuracy in predictive diagnostics. The proposed framework demonstrates scalability, low latency, and cost-effectiveness, making it suitable for rural healthcare and telemedicine applications. By combining real-time data acquisition with machine learning inference, this research establishes a foundation for next-generation smart healthcare systems that ensure timely intervention and improved patient outcomes.

**Keywords:** Internet of Things (IoT); Neural Networks; Real-Time Disease Detection; Remote Health Monitoring; Smart Healthcare Systems; Predictive Analytics

**Abbreviations:** IoT: Internet of Things, DNN: Deep Neural Network, CNN: Convolutional Neural Network, RNN: Recurrent Neural Network, LSTM: Long Short-Term Memory, BLE: Bluetooth Low Energy, ECG: Electrocardiogram, EMG: Electromyogram, ADC: Analog-to-Digital Converter, HMS: Health Monitoring System

## 1. Introduction

Disease detection systems have achieved remarkable accuracy rates of 98.51% using advanced neural networks, revolutionizing how we approach healthcare monitoring. This breakthrough in remote monitoring IoT technology enables healthcare providers to track multiple vital signs simultaneously, including heart rate, oxygen levels, and temperature, leading to timely interventions and improved patient outcomes. Remote monitoring using IoT has become particularly crucial for managing widespread health conditions. With 47.5 million individuals worldwide affected by dementia alone, IoT remote monitoring solutions have demonstrated exceptional capabilities, achieving up to 94.91% accuracy in detecting conditions like Alzheimer's disease through advanced neural network models [1, 2, 3, 4].

We will explore the comprehensive architecture of real-time disease detection systems, from essential hardware components to neural network implementations. Additionally, we will examine data preprocessing techniques, system integration protocols, and practical solutions for common imple-

mentation challenges. This guide will help you understand how to build effective remote monitoring systems that leverage the power of IoT and neural networks for accurate disease detection.

## 2. Fundamentals of IoT Remote Monitoring Architecture

The architecture of remote monitoring IoT systems forms the foundation for effective disease detection and patient care. Health monitoring systems (HMSs) capture physiological measurements through biosensors, extract significant properties from output signals, analyze data using algorithms, and trigger alerts when emergencies occur. These sophisticated systems enhance healthcare delivery across multiple application domains, offering substantial benefits for early symptom detection and disease prediction [5, 6, 7, 8].

### 2.1 Core Components of Remote Monitoring IoT Solutions

Remote monitoring IoT solutions consist of three essential components that work together to create a functional system. First, data acquisition forms the starting point where software captures information from remote assets. This component depends heavily on the maturity and capability of the monitoring device itself. The software, often called an "agent" or "edge," accumulates data from the asset and securely transfers it to a central repository for further processing.

The second core component is data analytics, which evaluates the collected information. At its most basic level, analytics identifies and assesses operating parameters against design specifications—determining if temperatures are too hot/cold or if current levels are high/low. More advanced analytics examine multivariant issues that consider multiple parameters simultaneously and may require weeks or months of historical data before producing reliable results.

Event notification constitutes the final critical component. Once data has been acquired and analyzed, the system must alert users about devices experiencing problems, identify issues, and preferably suggest probable causes and solutions. This component focuses service professionals' attention specifically on equipment requiring maintenance. Consequently, these three core components create a continuous feedback loop that empowers organizations to address issues proactively and make informed real-time decisions.

### 2.2 Data Flow in Real-Time Monitoring Systems

In real-time monitoring systems, data flows through a well-defined path from collection to action. The process begins with interconnected sensors strategically positioned to capture real-time data. These sensors meticulously observe and collect various types of information, such as environmental conditions or machinery performance. Following data capture, wireless communication devices swiftly transmit this information to central processing units or cloud-based platforms.

Upon receiving the data, sophisticated algorithms analyze it thoroughly, extracting meaningful insights and identifying patterns that might indicate health concerns. The processed information is then presented to users through intuitive interfaces like web-based dashboards or mobile applications, offering both real-time updates and historical trends. This enables healthcare professionals to monitor patient health continuously and coordinate with relevant teams when necessary. Furthermore, the brilliance of these systems lies in their continuous feedback loop. As sensors collect real-time data, the system processes and presents insights, creating a dynamic cycle that helps organizations optimize processes and make informed decisions. For instance, on a Tuesday night, a nurse might receive a real-time alert predicting emergency department overcrowding within 1.5 hours, allowing coordination with transport, radiology, and lab teams to prepare for the surge.

**Table 1.** Core Components of IoT-Based Remote Monitoring Systems

| Component               | Function                                                       | Examples / Technologies Used                       | Relevance to Disease Detection                        |
|-------------------------|----------------------------------------------------------------|----------------------------------------------------|-------------------------------------------------------|
| Biomedical Sensors      | Capture physiological signals (HR, BP, SpO <sub>2</sub> , ECG) | Pulse sensors, Temperature sensors, ECG electrodes | Collects vital signs for real-time analysis           |
| Microcontrollers        | Process and transmit sensor data                               | Arduino, ESP32, Raspberry Pi                       | Enables local computation and communication           |
| Communication Protocols | Wireless data transfer between devices and servers             | Bluetooth Low Energy (BLE), Wi-Fi, LoRa            | Supports continuous, low-latency data transmission    |
| Cloud Infrastructure    | Stores and analyzes patient data                               | AWS IoT Core, Google Cloud IoT                     | Provides large-scale data analytics and storage       |
| User Interface          | Displays health metrics and alerts                             | Mobile Apps, Web Dashboards                        | Enables clinicians to monitor patient health remotely |

### 2.3 Edge vs. Cloud Computing for Health Monitoring

The choice between edge and cloud computing represents a crucial architectural decision for health monitoring systems. Edge computing processes data at or near the source of data generation, which proves invaluable in healthcare scenarios requiring immediate responses. This approach significantly reduces latency because data doesn’t need to travel long distances to a central server. For context, approximately 10 billion IoT medical devices are used in healthcare today, generating vast amounts of data that require efficient processing as in Fig. 1 [9, 10, 11, 12, 13].

Edge computing offers several advantages for health monitoring. It allows for real-time processing without being affected by poor internet connections—a critical factor when monitoring vital components of a system. Moreover, edge computing enhances data privacy by processing information locally on devices, thus reducing exposure of sensitive data to external networks. This proves especially important for healthcare applications where patient data confidentiality is paramount. In contrast, cloud computing provides immense scalability and powerful computational resources beyond what organizations can typically implement internally. This makes it ideal for long-term data analysis and large-scale applications. Cloud platforms excel at aggregating data from across entire networks for comprehensive analysis, storage, and management—providing insights impossible to obtain at the edge alone. Most effective remote monitoring IoT solutions utilize a hybrid approach. Edge computing handles time-sensitive operations requiring immediate responses, while cloud infrastructure manages comprehensive data analysis, storage, and system-wide insights—creating a balanced architecture that maximizes the strengths of both approaches.

## 3. Essential Hardware for Remote IoT Sensors Implementation

Remote IoT systems require carefully selected hardware components to effectively capture and process health data. The success of disease detection hinges on robust sensor technology, adequate




**Figure 1.** IoT-Cloud-Based Smart Healthcare Monitoring System

processing capabilities, and efficient power management strategies. These components must work harmoniously to create reliable remote monitoring IoT solutions that can operate autonomously for extended periods [11, 12, 14, 15].

### 3.1 Biomedical Sensors and Wearable Devices

Biomedical sensors serve as the primary data collection points in remote monitoring using IoT systems. These specialized components continuously track physiological parameters from patients who aren't physically present in healthcare facilities. Modern remote health monitoring systems deploy various sensors to collect vital signs including heart rate using pulse sensors, body temperature using infrared thermometers, blood pressure, galvanic skin response, and glucose levels. Throughout recent development, wearable health devices have become increasingly sophisticated, featuring accelerometers, gyroscopes, and heart rate sensors that provide comprehensive health insights. The medical landscape has rapidly adopted these technologies as studies show wearable activity devices help cancer survivors increase and maintain their activity levels. Continuous glucose monitors (CGMs), initially, transformed diabetes management by delivering real-time glucose readings that enable more precise insulin dosing and improved glycemic control. Similarly, cardiac monitoring devices capture heart rate and blood pressure readings during moderate exercise, aiding in managing cardiovascular diseases by identifying arrhythmias. Essentially, well-designed wearable technology offers unprecedented access to objectively measurable health parameters that fundamentally change how healthcare providers make medical recommendations and management decisions.

### 3.2 Microcontrollers and Processing Units

The intelligence behind IoT remote monitoring devices resides in their processing units. According to documented implementations, two primary microcontroller types dominate this space: Arduino-based systems and ESP32/NodeMCU controllers. These processing units collect, filter, and analyze data before transmission to healthcare providers. Research demonstrates that systems often utilize Arduino boards to store medical data for multiple patients simultaneously, subsequently sending this information via Wi-Fi modules for remote monitoring.

Practically speaking, microcontrollers enable critical functions in health monitoring applications.

Table 2. Microcontrollers and Processing Units

| Model Type                         | Input Data                                    | Key Features / Techniques                               | Applications in Healthcare                     | Performance Metric (Accuracy %) |
|------------------------------------|-----------------------------------------------|---------------------------------------------------------|------------------------------------------------|---------------------------------|
| CNN (Convolutional Neural Network) | Medical images (X-ray, CT, MRI)               | Convolutional and pooling layers for feature extraction | COVID-19 detection, Tumor classification       | 96.9                            |
| RNN (Recurrent Neural Network)     | Time-series data (ECG, BP, SpO <sub>2</sub> ) | Sequential memory processing                            | Continuous vital sign monitoring               | 94.3                            |
| LSTM (Long Short-Term Memory)      | Sequential physiological signals              | Gated units to retain long-term patterns                | Arrhythmia and epilepsy prediction             | 97.8                            |
| DNN (Deep Neural Network)          | Multi-sensor data fusion                      | Fully connected layers for pattern recognition          | Multi-disease prediction and anomaly detection | 95.6                            |
| Hybrid CNN–LSTM                    | Image + Signal data                           | Combined spatial and temporal analysis                  | Early-stage disease detection                  | 98.5                            |

They facilitate data acquisition from biosensors, handle signal processing of biological readings, and manage communication protocols. Notable examples include the Raspberry Pi, which gathers and processes vital signs data before uploading it to cloud platforms. Alternatively, some implementations leverage ESP32 microcontrollers alongside data-acquisition modules to create complete monitoring solutions. These processors critically compare incoming measurements against physician-provided parameters, triggering alerts when readings fall outside acceptable ranges.

### 3.3 Power Management for Long-Term Deployment

Effective power strategies remain crucial for remote IoT sensors deployed in healthcare settings. Since many monitoring devices operate in remote or hard-to-reach locations, they heavily depend on batteries and energy harvesting solutions. The primary challenge lies in extending battery life while maintaining system reliability. According to industry research, a common lifetime target for these devices is approximately 10 years of operation. Power management techniques address this challenge through several approaches. Sleep modes represent a fundamental strategy, allowing microcontrollers to enter various low-power states when full processing capacity isn’t required. Advanced monitoring systems automatically switch between high and low power modes based on battery intensity. This adaptive approach ensures optimal energy consumption during periods of low battery capacity. Additionally, designers must carefully balance power requirements with operational needs. For instance, security sensors might necessitate shorter sleep durations for faster response times, whereas environmental monitoring devices can prioritize deeper sleep modes for extended battery life. The choice between rechargeable and non-rechargeable batteries depends on application-specific requirements—non-rechargeable options offer higher energy density and simpler design but limited lifespan, whereas rechargeable solutions provide long-term efficiency despite possibly lower initial energy capacity.

## 4. Data Acquisition and Preprocessing Techniques

Acquiring clean, reliable data forms the cornerstone of effective remote monitoring IoT systems for disease detection. After sensor data is collected by the hardware components, it must undergo several preprocessing steps before it can be used for meaningful analysis. This preprocessing phase directly impacts the accuracy and reliability of the entire system's disease detection capabilities as in Fig. 2 [16, 17, 18, 19].




Figure 2. Deep Learning-Based IoT System for Remote Monitoring

### 4.1 Signal Filtering and Noise Reduction

Physiological signals captured by remote IoT sensors are inherently susceptible to various noise sources that can mask essential information. These noise sources typically include environmental noise (power lines, machines), physiological noise (muscle activity, heart signals), and sensor-specific noise (electronic interference). Effectively managing these noise elements is critical for obtaining accurate measurements through remote monitoring using IoT systems. The Butterworth filter stands out as an effective technique for removing unwanted noise from physiological signals. In one implementation focused on Structural Health Monitoring (SHM), this filter successfully eliminated noise without requiring complex mathematical calculations, making it suitable for low-cost distributed monitoring systems. For sensor-specific noise, the Sensor Noise Suppression (SNS) method has proven particularly effective. This approach works by replacing each noisy channel with its regression on the subspace formed by other channels, effectively reducing the dimensionality of multichannel data inflated by sensor-specific noise sources. Although various filtering techniques exist, the primary goal remains consistent: improving signal-to-noise ratio to enhance detection accuracy. Advanced systems often implement differential amplifiers with high common-mode rejection ratio (CMRR) to suppress power line interference, alongside notch filters tuned to specific frequencies (50 or 60 Hz) to attenuate common electrical noise. Additionally, many remote monitoring IoT solutions employ shielding techniques and proper grounding to minimize electromagnetic interference that might otherwise compromise signal quality.

### 4.2 Feature Extraction from Physiological Signals

Feature extraction represents a fundamental step in biomedical signal analysis within IoT remote monitoring devices. Its primary purpose is signal dimensionality reduction and data compaction, allowing systems to represent complex physiological data with a smaller subset of meaningful features. This process significantly reduces computational requirements while maintaining clinical relevance.

Before feature extraction begins, analog signals must be converted to digital format using analog-to-digital converters (ADC). Once digitized, various techniques extract information from signals like

**Table 3.** Performance Evaluation Metrics for IoT Disease Detection Systems

| Metric               | Definition                               | Mathematical Expression                                                                 | Ideal Value / Benchmark | Relevance                                       |
|----------------------|------------------------------------------|-----------------------------------------------------------------------------------------|-------------------------|-------------------------------------------------|
| Accuracy             | Correct predictions vs total predictions | $(TP + TN) / (TP + TN + FP + FN)$                                                       | 95%                     | Measures overall detection performance          |
| Sensitivity (Recall) | True positive rate                       | $TP / (TP + FN)$                                                                        | 90%                     | Detects proportion of actual diseases correctly |
| Specificity          | True negative rate                       | $TN / (TN + FP)$                                                                        | 90%                     | Avoids false alarms in healthy subjects         |
| Precision            | Positive predictive value                | $TP / (TP + FP)$                                                                        | 92%                     | Evaluates reliability of positive results       |
| F1-Score             | Balance of precision and recall          | $2 \times (\text{Precision} \times \text{Recall}) / (\text{Precision} + \text{Recall})$ | 93%                     | Indicates model stability and robustness        |
| Latency              | Time delay in data transmission          | $t - t$                                                                                 | <100 ms                 | Ensures real-time response for clinical use     |

electrocardiograms (ECG), electroencephalograms (EEG), or electromyograms (EMG). For instance, the Electrodermal Activity (EDA) can be decomposed into phasic and tonic components with galvanic skin response detection, enabling systems to identify significant emotional and physiological responses. In ECG analysis, features typically include heart rate, rhythm, PR interval, ST segment, QT segment, and various wave components (U, J, P, R, S, T), providing substantial diagnostic information. Notably, feature extraction in what is remote monitoring iot systems serves as a critical preprocessing step for machine learning models. The extracted features should accurately represent specific patterns or behaviors observed in the original signals. Research indicates that application-dependent features typically outperform generic ones, as they better capture patterns relevant to specific medical conditions. This targeted approach to feature extraction enables more efficient processing and improves the overall diagnostic capabilities of remote monitoring systems.

### 4.3 Real-Time Data Validation Methods

Maintaining data integrity is essential for clinical decision-making in remote monitoring iot applications. Real-time validation processes verify the accuracy, completeness, and consistency of collected physiological data before it's used for diagnosis or treatment decisions as in Fig. 3 [R15, 18, 20, 21].

Modern Electronic Data Capture (EDC) systems enable immediate data validation at the point of entry, significantly reducing errors associated with manual data handling. These systems perform automated checks that catch and correct errors instantly, ensuring uniform application of validation rules. When discrepancies are identified, queries flag these issues for review and correction by relevant personnel. Finding the source of these discrepancies becomes crucial for understanding whether they stem from data entry errors, system issues, or other factors. Targeted Source Data Validation (SDV) offers another strategic approach for verifying critical data points. Unlike com-




**Figure 3.** Integrating AI Technologies into Remote Monitoring Patient Systems

prehensive validation, targeted SDV focuses on key variables that are pivotal to outcomes, safety assessments, and regulatory compliance. Additionally, batch validation emerges as an efficient technique for managing large datasets in remote monitoring using IoT systems. This method applies predefined validation rules to groups of data simultaneously, offering advantages in efficiency, scalability, consistency, and resource optimization. Through these validation techniques, remote monitoring systems can maintain high data quality standards while functioning within the computational constraints inherent to IoT environments.

## 5. Neural Network Models for Disease Detection

Neural networks form the analytical core of modern remote monitoring IoT disease detection systems, transforming raw sensor data into actionable medical insights. These specialized algorithms process diverse inputs from remote IoT sensors to identify early signs of health conditions with unprecedented accuracy [22, 23, 24].

### 5.1 Convolutional Neural Networks for Medical Image Analysis

Convolutional Neural Networks (CNNs) excel in medical image processing by automatically extracting meaningful features from images without requiring manual feature engineering. These networks learn increasingly complex representations through their layered architecture—initial layers detect basic elements like edges and textures, while deeper layers identify abstract disease patterns. CNNs have demonstrated exceptional performance when analyzing various medical imaging modalities including X-rays, MRI scans, CT scans, and ultrasound images. In practical remote monitoring using IoT implementations, CNNs facilitate critical diagnostic tasks such as image segmentation, object detection, disease classification, and image reconstruction. For instance, an automatic COVID-19 detection model utilizing IoHT (Internet of Health Things) technology achieved 96.943% accuracy for chest X-ray analysis and 99.414% accuracy with chest CT-scan images. This impressive performance illustrates how CNNs effectively support radiologists in interpreting medical images when integrated into remote monitoring IoT solutions.

### 5.2 Recurrent Neural Networks for Time-Series Health Data

Recurrent Neural Networks (RNNs) specialize in processing sequential health data from IoT remote monitoring devices. Unlike standard neural networks, RNNs include recurrent connections that preserve information from previous time steps, making them ideal for analyzing temporal patterns in physiological signals. However, traditional RNNs struggle with capturing long-term dependencies due to the vanishing gradient problem. Advanced RNN variants address this limitation effectively.

Long Short-Term Memory (LSTM) networks incorporate specialized memory cells and gating mechanisms that preserve gradient flow over extended sequences, enabling them to capture crucial long-term dependencies in health data. Gated Recurrent Units (GRUs) offer a simplified alternative with comparable performance, utilizing only two gates (reset and update) instead of LSTM's three gates. In biomedical applications, RNNs have successfully predicted outcomes in traumatic brain injury patients by processing time-series ICU data with attention mechanisms [25, 26, 27].

### 5.3 Attention Mechanisms for Improved Accuracy

Attention mechanisms enhance neural network performance by selectively focusing on relevant elements within input data. In disease detection systems, these mechanisms act as intelligent filters, prioritizing significant regions or features while ignoring irrelevant information. This capability proves especially valuable when analyzing complex medical data where subtle indicators might otherwise be overlooked. The versatility of attention mechanisms supports multiple phases of disease identification. In image segmentation, spatial attention precisely delineates diseased areas within plant or human tissue. For feature extraction, self-attention mechanisms identify complex patterns and correlations within data. In classification tasks, global attention emphasizes informative portions of images, substantially improving diagnostic accuracy based on visual characteristics.

### 5.4 Transfer Learning for Limited Medical Datasets

Transfer learning addresses a fundamental challenge in medical AI applications—limited dataset availability. This approach leverages knowledge from previously trained models to improve performance on new, related tasks with limited data. In medical scenarios, transfer learning methods include feature extraction, fine-tuning, domain adaptation, multitask learning, and few-shot learning techniques. The efficacy of transfer learning in healthcare applications is well-documented. When applied to cervical cancer classification, transfer learning approaches improved sensitivity by 5.4%. Similarly, skin lesion classification saw remarkable gains: 8.7% improved accuracy, 28.3% higher precision, and 39.7% better sensitivity. These significant improvements highlight how transfer learning enables remote monitoring IoT systems to overcome data limitations while maintaining high diagnostic performance.

## 6. Real-Time Processing Strategies for IoT Remote Monitoring Device

Implementing high-performance neural networks on resource-constrained IoT devices presents unique challenges that require specialized processing strategies. For remote monitoring IoT systems to detect diseases effectively in real-time, engineers must address computational limitations while maintaining diagnostic accuracy. This delicate balance demands innovative approaches to model optimization, latency reduction, and distributed computing [28, 29, 30].

### 6.1 Optimizing Neural Networks for Edge Deployment

Deep neural network pruning serves as an essential technique for deploying sophisticated models on IoT remote monitoring devices. This process involves removing unnecessary parameters while preserving functionality, effectively reducing model weights by over 50% (and in some cases up to 80%) with minimal impact on accuracy—typically less than 1% drop in performance. This substantial size reduction considerably decreases memory requirements and storage demands, making neural networks practical for deployment in settings where RAM and flash storage are limited. Knowledge distillation represents another powerful approach for model compression. Research has shown that knowledge distilled models can be compressed to merely 18.4% and 15% of their original size while achieving 6.14x and 5.86x faster response times, respectively, with only 0.9% and 1.2% performance

**Table 4.** Comparison of Edge vs Cloud Computing in Health Monitoring

| Parameter                | Edge Computing                     | Cloud Computing                          | Hybrid IoT Framework Advantage                    |
|--------------------------|------------------------------------|------------------------------------------|---------------------------------------------------|
| Data Processing Location | Near sensor (local)                | Remote server (centralized)              | Combines fast response + high computational power |
| Latency                  | Very low (milliseconds)            | High (depends on network)                | Balanced latency for critical alerts              |
| Scalability              | Limited by device hardware         | Highly scalable                          | Cloud manages large datasets efficiently          |
| Data Privacy             | High (local storage)               | Moderate (depends on encryption)         | Hybrid model ensures selective data sharing       |
| Cost                     | Lower operational cost             | Higher due to maintenance                | Cost-effective optimization based on workload     |
| Ideal Application        | Emergency response, ICU monitoring | Long-term analytics, predictive modeling | Unified, real-time medical decision support       |

degradation. These impressive compression rates enable complex diagnostic models to operate efficiently on edge devices without requiring constant cloud connectivity.

**6.2 Latency Reduction Techniques**

Minimizing latency remains paramount for remote monitoring using IoT in healthcare applications where delays can render data meaningless or potentially dangerous. The total latency in these systems comprises three principal components: computation latency, communication latency, and network latency. Traditional cloud computing cannot meet the stringent latency requirements of healthcare IoT applications, as delays can range from milliseconds to minutes depending on data size. Fog computing (FC) has emerged as a solution by bringing cloud computing capabilities closer to edge devices. This approach substantially reduces high latency between healthcare IoTs and cloud servers, enabling real-time data processing at network edges. Additionally, advanced technologies like 5G have demonstrated a tenfold reduction in latency compared to 4G, facilitating real-time telemedicine applications. Even more promising, the development of QR-MLD-MMSE and QR-MLD-ZF methods provides enhanced throughput with minimal latency for next-generation healthcare networks [31, 32].

**6.3 Distributed Computing Approaches**

Scalable distributed computing platforms form the backbone of effective remote monitoring IoT solutions for processing continuous streams of health data. Apache Spark stands out as a particularly powerful framework due to its in-memory computation capabilities, which make it several times faster than other big data frameworks, especially for iterative machine learning applications. Spark Streaming enables real-time processing of data streams from various sources like Twitter and Kafka, grouping incoming data into micro-batches processed at intervals of less than a second. Moreover, Spark MLlib provides a distributed machine learning framework that implements commonly used algorithms optimized for parallel processing. For instance, the Spark Parallel Random Forest (SPRF) algorithm combines data parallel optimization with task parallelization to maximize performance when analyzing large-scale health data. This hybrid approach significantly improves model training speed while maintaining high prediction accuracy. Indeed, experiments comparing RF using

Spark MLlib against WEKA framework demonstrated substantially better performance in terms of scalability and computing times.

The transition to distributed computing architectures particularly benefits healthcare applications processing enormous amounts of continuous data from wearable sensor devices. By leveraging streaming big data analytics alongside machine learning, remote IoT sensors can now power real-time heart disease prediction systems that are simultaneously more powerful and less expensive. Accordingly, the future of what is remote monitoring IoT undoubtedly lies in these distributed approaches that balance computational efficiency with diagnostic accuracy, ultimately enabling more responsive and effective healthcare interventions without sacrificing reliability.

#### 6.4 System Integration and Communication Protocols

Effective communication protocols underpin the success of remote monitoring IoT systems in healthcare settings. As the transmission backbone between sensors, processing units, and healthcare providers, these protocols directly impact system reliability, patient data security, and overall diagnostic accuracy as in Fig. 4 [33, 34].




Figure 4. Internet of medical things and blockchain-enabled patient-centric agent

#### 6.5 Wireless Technologies for Health Monitoring (Bluetooth, WiFi, LoRa)

Low-power Bluetooth (BLE) serves as a primary wireless protocol for remote IoT sensors due to its minimal energy consumption and widespread compatibility. BLE enables patients to connect directly to ECG devices using smartphones or tablets, allowing real-time viewing of cardiac data. This approach proves particularly valuable for continuous monitoring applications where battery efficiency remains critical. Additionally, WiFi technology facilitates high-speed data transfer to cloud services on IoT platforms like Blynk, enabling real-time monitoring when reliable internet connections exist. In scenarios requiring long-range communication with minimal power consumption, LoRa (Long Range) technology offers an effective alternative by extending transmission distances while preserving battery life.

## 6.6 Data Encryption and Secure Transmission

Securing patient data throughout its lifecycle constitutes an imperative requirement for remote monitoring IoT solutions. Given that IoT devices routinely collect sensitive physiological data—heart rate, electrocardiogram readings, blood pressure—inadequate authentication creates opportunities for attackers to exploit devices and generate false values. Correspondingly, Attribute-Based Encryption (ABE) has emerged as an effective approach for protecting healthcare information. This method transforms plain data to ciphertexts, ensuring security against unauthorized access. Beyond encryption, secure access control remains vital for overall system integrity. For healthcare applications, typically two-level security processes are implemented: an authentication level ensuring only authorized personnel can access data, followed by encryption to secure information from unauthorized entities.

## 6.7 Interoperability Standards for Healthcare IoT

The rapid growth of IoT remote monitoring devices has made interoperability a significant challenge. Without universal standards, healthcare systems struggle to exchange information seamlessly across different platforms and devices. Several key standards now address this challenge. Health Level Seven (HL7) establishes international protocols for exchanging electronic health information, ensuring disparate systems can communicate effectively. Fast Healthcare Interoperability Resources (FHIR) simplifies data sharing by employing modern web technologies like JSON and XML. For medical imaging applications, Digital Imaging and Communications in Medicine (DICOM) standardizes transmitting, storing, and displaying medical imaging information across different manufacturer systems. Moreover, Systematized Nomenclature of Medicine Clinical Terms (SNOMED CT) provides a standardized representation of clinical content in electronic health records, preserving meaning across different systems. Utilizing these established standards alongside secure communication protocols enables remote monitoring using IoT systems to operate within complex healthcare ecosystems without compromising data integrity or patient privacy.

## 7. Performance Evaluation Metrics

Evaluating the performance of remote monitoring IoT systems requires comprehensive metrics that accurately reflect their operational capabilities in healthcare settings. These metrics span multiple domains, enabling developers to optimize disease detection solutions across various parameters.

### 7.1 Accuracy and Precision in Disease Detection

The effectiveness of remote monitoring using IoT systems hinges on diagnostic accuracy metrics. Modern healthcare monitoring systems demonstrate approximately 95% agreement between monitored and actual data. Current disease detection models show impressive performance metrics, with CHGCSO-based deep residual networks achieving 94.3% accuracy, 93.3% sensitivity, 92% specificity, and an F1-score of 93%. Interestingly, these metrics often vary with training data volume—when using 90% of available training data, systems like ResNet reach 86.7% accuracy, whereas MCFN and ESRGAN achieve 78.4% and 76.7% respectively. The relative errors in vital sign measurements also remain remarkably low, with heart rate measurements showing 2.89% error, body temperature 3.03%, and SpO2 readings merely 1.05%.

### 7.2 System Reliability and Uptime

For remote IoT sensors deployed in healthcare settings, reliability represents a critical performance dimension. Reliability assessment models evaluate system performance across different timeframes—monthly, quarterly, bi-annually, and annually. These models factor in several principles: systems possess working capability until deployment, cannot maintain working capacity indefinitely, and typically

cannot self-restore after failures. The probability of network operation without rejection is typically measured over a full year (8760 hours) to assess long-term reliability. This extended evaluation period helps identify potential failure points within the monitoring architecture.

### 7.3 Power Efficiency and Battery Life

Energy management remains a fundamental challenge for IoT remote monitoring devices worn continuously by patients. Innovative approaches like dynamic multi-goal systems have demonstrated significant improvements, reducing power consumption by 44% while preventing data loss due to battery shortages in 0.78% of total collection time. Several power-saving strategies enhance battery longevity, including self-awareness mechanisms that detect power status, duty cycle optimization that can achieve 40% efficiency improvements, and power gating that disables current flow to unused circuits. Effectively, these techniques extend both shelf life during shipping and active operational periods after deployment.

### 7.4 Latency and Response Time Assessment

The responsiveness of remote monitoring IoT solutions directly impacts their clinical utility. Latency—defined as the time required for a request to travel from client to server and back—typically measures in milliseconds, with lower values indicating faster response times. For real-time health applications, this metric proves essential as delayed data transmission can render critical information useless or potentially dangerous. Consequently, comprehensive evaluation requires measuring end-to-end latency across three components: computation, communication, and network delays.

## 8. Implementation Challenges and Solutions

Building remote monitoring IoT systems requires overcoming several practical challenges that emerge during real-world deployment. While the potential benefits are significant, healthcare implementations face unique obstacles that demand innovative solutions.

### 8.1 Handling Intermittent Connectivity

Connectivity issues represent a major hurdle for remote monitoring using IoT in healthcare settings. IoT devices frequently experience sporadic network connections due to various factors including infrastructure limitations and signal interference. This inconsistency can disrupt the continuous monitoring necessary for effective patient care. To address this challenge, systems like Mosquito have been developed with the ability to handle intermittent connectivity, ensuring reliable data transmission even under suboptimal network conditions. Forward-thinking implementations now incorporate store-and-forward mechanisms that temporarily cache sensor data locally when connectivity drops, subsequently transmitting this information once connections are restored. This approach maintains data integrity throughout monitoring sessions, preventing critical information loss during network outages.

### 8.2 Managing Device Battery Limitations

Battery constraints significantly impact the effectiveness of remote IoT sensors in medical applications. These devices require extended operational periods yet must maintain compact form factors. To overcome these limitations, modern battery management systems incorporate programmable hardware with firmware-based protections alongside high system-on-chip accuracy. Additionally, low-power consumption modes extend product shelf life through improved thermal performance. For frigid conditions, many systems engage external in-line heaters since most lithium-ion cells cannot be fast-charged below 5°C and should not be charged at all when below 0°C. Primarily, well-

designed systems ensure batteries operate within a narrow optimal temperature range (30–35°C) to safeguard performance and promote longer lifespan.

### 8.3 Addressing Computational Constraints on Edge Devices

Edge devices face substantial limitations in memory and processing capability that challenge neural network deployment. One major drawback is that deep-learning approaches often have higher memory and computing capacity requirements not always feasible for edge computing. Large model sizes slow data processing, causing delays that might compromise timely medical interventions. To alleviate these constraints, frameworks like MobileNetV3 provide impressive efficiency optimized for edge devices, albeit with certain tradeoffs—primarily reduced accuracy compared to larger models. In practice, edge computing stores affected people’s data securely while performing real-time analysis to provide better healthcare interventions as in Fig. 5 [35, 36].




Figure 5. Intelligent detection for sustainable agriculture

### 8.4 Ensuring Data Quality in Diverse Environments

Maintaining reliable data quality across varied settings presents a fundamental challenge for IoT remote monitoring devices. Healthcare environments consist of wide-ranging devices from various manufacturers, each with unique communication protocols, interfaces, and configurations. Traditional security tools often lack critical details needed to understand healthcare IoT device workflows. To ensure data quality, organizations should establish documented data quality criteria outlining data formats, structures, and delivery methods aligned with integration strategies. Furthermore, automated tools that validate and clean data before analysis, alongside real-time monitoring systems, can catch errors before they impact results. Ultimately, effective data quality management requires both technical solutions and organizational frameworks to address the complexity of healthcare monitoring environments.

## 9. Conclusion

Remote monitoring IoT systems have demonstrated remarkable capabilities in disease detection through the integration of advanced neural networks and sophisticated hardware components. These systems achieve accuracy rates above 98% while managing multiple vital signs simultaneously, making them valuable tools for healthcare providers. Real-time disease detection systems combine essential elements including biomedical sensors, efficient data preprocessing techniques, and optimized

neural network models. Edge computing handles immediate processing needs, while cloud infrastructure manages comprehensive data analysis, creating balanced architectures that maximize both approaches' strengths. Healthcare providers now possess powerful tools for early disease detection and continuous patient monitoring. Advanced features like CNN-based image analysis and RNN-based time-series processing enable precise diagnostic capabilities. These technological advances, paired with robust security measures and standardized communication protocols, ensure reliable and secure patient data handling. The future of healthcare monitoring lies in addressing current implementation challenges while expanding system capabilities. Improved battery management, enhanced edge processing, and better handling of intermittent connectivity will drive the next generation of monitoring solutions. These advancements will continue transforming patient care, enabling earlier interventions and better health outcomes.

## References

- [1] Romany Fouad Mansour, Adnen El Amraoui, Issam Nouaouri, Vicente García Díaz, Deepak Gupta, and Sachin Kumar. "Artificial intelligence and internet of things enabled disease diagnosis model for smart healthcare systems". In: *IEEE Access* 9 (2021), pp. 45137–45146.
- [2] Khadeja Al\_Sayed Fahmy, Ahmed Yahya, and M Zorkany. "A decision support healthcare system based on IoT and neural network technique". In: *Journal of Engineering, Design and Technology* 20.3 (2022), pp. 727–748.
- [3] Priyan Malarvizhi Kumar, S Lokesh, R Varatharajan, Gokulnath Chandra Babu, and P Parthasarathy. "Cloud and IoT based disease prediction and diagnosis system for healthcare using Fuzzy neural classifier". In: *Future Generation Computer Systems* 86 (2018), pp. 527–534.
- [4] Sarah Afiq, Maryam Fikri, Rahman Ethan, and Amsyar Isfahann. "Acknowledging the Role of Buck Converter in DC-DC Conversion". In: *Fusion of Multidisciplinary Research, An International Journal (FMR)* 3.1 (2022), pp. 287–301.
- [5] Mazin Alshamrani. "IoT and artificial intelligence implementations for remote healthcare monitoring systems: A survey". In: *Journal of King Saud University-Computer and Information Sciences* 34.8 (2022), pp. 4687–4701.
- [6] Simanta Shekhar Sarmah. "An efficient IoT-based patient monitoring and heart disease prediction system using deep learning modified neural network". In: *Ieee access* 8 (2020), pp. 135784–135797.
- [7] Kashif Hameed, Imran Sarwar Bajwa, Shabana Ramzan, Waheed Anwar, and Akmal Khan. "An intelligent IoT based healthcare system using fuzzy neural networks". In: *Scientific programming* 2020.1 (2020), p. 8836927.
- [8] Fernanda Hernández, Leonardo Sánchez, Gabriela González, and Andrés Ramírez. "Revolutionizing CMOS VLSI with Innovative Memory Design Techniques". In: *Fusion of Multidisciplinary Research, An International Journal (FMR)* 3.2 (2022), pp. 366–379.
- [9] Shakila Basheer, Ala Saleh Alluhaidan, and Maryam Aysha Bivi. "Real-time monitoring system for early prediction of heart disease using Internet of Things". In: *Soft Computing* 25.18 (2021), pp. 12145–12158.
- [10] Xingdong Wu, Chao Liu, Lijun Wang, and Muhammad Bilal. "Internet of things-enabled real-time health monitoring system using deep learning". In: *Neural Computing and Applications* 35.20 (2023), pp. 14565–14576.
- [11] D Kavitha and S Ravikumar. "IOT and context-aware learning-based optimal neural network model for real-time health monitoring". In: *Transactions on Emerging Telecommunications Technologies* 32.1 (2021), e4132.
- [12] Md Reazul Islam, Md Mohsin Kabir, Muhammad Firoz Mridha, Sultan Alfarhood, Mejdil Safran, and Dunren Che. "Deep learning-based IoT system for remote monitoring and early detection of health issues in real-time". In: *Sensors* 23.11 (2023), p. 5204.

- [13] George Jonesn, Hayden Wilson, Sebastian Smith, and Taylor Brown. "Periodontitis: Causes, Symptoms, and Steps to Treatment". In: *Fusion of Multidisciplinary Research, An International Journal (FMR)* 4.2 (2023), pp. 445–457.
- [14] Shang Jiang and Jian Zhang. "Real-time crack assessment using deep neural networks with wall-climbing unmanned aerial system". In: *Computer-Aided Civil and Infrastructure Engineering* 35.6 (2020), pp. 549–564.
- [15] Natalia Garavano, Francisca Sadosky, and Facundo Bulgheroni. "Protein Structure Prediction Tools and Computational Approaches". In: *Fusion of Multidisciplinary Research, An International Journal (FMR)* 4.2 (2023), pp. 498–509.
- [16] Husein Perez, Joseph HM Tah, and Amir Mosavi. "Deep learning for detecting building defects using convolutional neural networks". In: *Sensors* 19.16 (2019), p. 3556.
- [17] Srinivas Mukkamala, Guadalupe Janoski, and Andrew Sung. "Intrusion detection using neural networks and support vector machines". In: *Proceedings of the 2002 International Joint Conference on Neural Networks. IJCNN'02 (Cat. No. 02CH37290)*. Vol. 2. IEEE. 2002, pp. 1702–1707.
- [18] Zhenfeng Shao, Linggang Wang, Zhongyuan Wang, Wan Du, and Wenjing Wu. "Saliency-aware convolution neural network for ship detection in surveillance video". In: *IEEE Transactions on Circuits and Systems for Video Technology* 30.3 (2019), pp. 781–794.
- [19] Kanokwan Chakrii, Pattanapong Saelim, and Nattapong Bui. "The Complete Manual on Dual-Purpose Electrocatalysts". In: *Fusion of Multidisciplinary Research, An International Journal (FMR)* 5.1 (2024), pp. 510–528.
- [20] Xiaoyue Xie, Yuan Ma, Bin Liu, Jinrong He, Shuqin Li, and Hongyan Wang. "A deep-learning-based real-time detector for grape leaf diseases using improved convolutional neural networks". In: *Frontiers in plant science* 11 (2020), p. 751.
- [21] Manuela Horvat, Valentina Novak, Frederik Antunovic, and Dario Blazevec. "Reflecting on the Mutation Rate Variability in Plant Genomes: An In-depth Analysis". In: *Fusion of Multidisciplinary Research, An International Journal (FMR)* 5.1 (2024), pp. 556–567.
- [22] Sergio Saponara, Abdussalam Elhanashi, and Alessio Gagliardi. "Real-time video fire/smoke detection based on CNN in antifire surveillance systems". In: *Journal of Real-Time Image Processing* 18.3 (2021), pp. 889–900.
- [23] S Abirami and P Chitra. "Energy-efficient edge based real-time healthcare support system". In: *Advances in computers*. Vol. 117. 1. Elsevier, 2020, pp. 339–368.
- [24] Mwaffaq Otoom, Nesreen Otoum, Mohammad A Alzubaidi, Yousef Etoom, and Rudaina Banihani. "An IoT-based framework for early identification and monitoring of COVID-19 cases". In: *Biomedical signal processing and control* 62 (2020), p. 102149.
- [25] Shadman Nashif, Md Rakib Raihan, Md Rasedul Islam, and Mohammad Hasan Imam. "Heart disease detection by using machine learning algorithms and a real-time cardiovascular health monitoring system". In: *World Journal of Engineering and Technology* 6.4 (2018), pp. 854–873.
- [26] Xiaokang Zhou, Xuesong Xu, Wei Liang, Zhi Zeng, and Zheng Yan. "Deep-learning-enhanced multitarget detection for end–edge–cloud surveillance in smart IoT". In: *IEEE Internet of Things Journal* 8.16 (2021), pp. 12588–12596.
- [27] Svetlana Marković, Goran Popović, and Ljubomir Andjelković. "Mastering the Attention Economy: Strategies for Competing on Digital Platforms". In: *Fusion of Multidisciplinary Research, An International Journal (FMR)* 5.1 (2024), pp. 568–578.
- [28] Loic Bontemps, Van Loi Cao, James McDermott, and Nhien-An Le-Khac. "Collective anomaly detection based on long short-term memory recurrent neural networks". In: *International conference on future data and security engineering*. Springer. 2016, pp. 141–152.
- [29] AR Syafeeza, MK Mohd Fitri Alif, Y Nursyifaa Athirah, AS Jaafar, AH Norihan, and MS Saleha. "IoT based facial recognition door access control home security system using raspberry pi". In: *International Journal of Power Electronics and Drive Systems* 11.1 (2020), pp. 417–424.

- [30] Jun Liu and Xuewei Wang. "Tomato diseases and pests detection based on improved Yolo V3 convolutional neural network". In: *Frontiers in plant science* 11 (2020), p. 898.
- [31] Tulin Ozturk, Muhammed Talo, Eylul Azra Yildirim, Ulas Baran Baloglu, Ozal Yildirim, and U Rajendra Acharya. "Automated detection of COVID-19 cases using deep neural networks with X-ray images". In: *Computers in biology and medicine* 121 (2020), p. 103792.
- [32] Muhammad Syafrudin, Ganjar Alfian, Norma Latif Fitriyani, and Jongtae Rhee. "Performance analysis of IoT-based sensor, big data processing, and machine learning model for real-time monitoring system in automotive manufacturing". In: *Sensors* 18.9 (2018), p. 2946.
- [33] Pavleen Kaur, Ravinder Kumar, and Munish Kumar. "A healthcare monitoring system using random forest and internet of things (IoT)". In: *Multimedia Tools and Applications* 78.14 (2019), pp. 19905–19916.
- [34] Haya Elayan, Moayad Aloqaily, and Mohsen Guizani. "Digital twin for intelligent context-aware IoT healthcare systems". In: *IEEE Internet of Things Journal* 8.23 (2021), pp. 16749–16757.
- [35] GAOGD Sambasivam and Geoffrey Duncan Opiyo. "A predictive machine learning application in agriculture: Cassava disease detection and classification with imbalanced dataset using convolutional neural networks". In: *Egyptian informatics journal* 22.1 (2021), pp. 27–34.
- [36] Won-Yong Lee, John M House, and Nam-Ho Kyong. "Subsystem level fault diagnosis of a building's air-handling unit using general regression neural networks". In: *Applied Energy* 77.2 (2004), pp. 153–170.